

RHODE ISLAND RADIATION CONTROL AGENCY
DENTAL X-RAY FACILITY INSPECTION REPORT

Inspection Report #: _____ **Registration** _____

Registrant (Name and Address):

Registrant Contact: _____ **Phone No.:** _____

Last Inspection Date: _____

Notice of Inspection Date: _____

Inspection Date: _____

Inspection Type:

- Initial
- Routine
- Reactive

Summary of Findings &

- No Items of Noncompliance, Clear RCA-9 issued
- Noncompliance, RCA-9 issued
- Noncompliance (Appendix A)
- Action on Previous Noncompliance (Appendix B)
- RCA Action

Inspector: _____ (signature) _____ (date signed)

Approved By: _____ (signature) _____ (date signed)

1) FACILITY ORGANIZATION:

Facility Supervisor: _____

Radiation Protection Person: _____

Dentist(s) responsible for interpretation: _____

Number of Radiographers: _____

Individuals operating X-ray equipment possess a current State license or course certificate and are adequately instructed in the safe operating procedures of the X-ray equipment [4.3.3(A)]: YES NO

2) X-RAY TUBE INVENTORY:

Tube Code ¹ :	20	21	22	23	24
--------------------------	----	----	----	----	----

Active: _____

Storage: _____

Total # of tubes: _____

Facility has shielding plan/evaluation conducted by a RI-registered Radiation Physics Service provider for all units in active use other than intraoral or closed-beam panoramic/cephalometric YES NO NA

3) OCCUPATIONAL RADIATION EXPOSURE

Personal dosimetry supplied to each individual expected to be exposed to radiation [1.10.3(A)] YES NO

Personal dosimetry devices exchanged at the appropriate frequency and processed by NVLAP approved and accredited processor [1.10.3(A)] YES NO

Personnel monitoring equipment properly worn [1.10.3(A)] YES NO

All personal dosimetry reports are reviewed and signed by the RSO when they are received [1.10.3(A)] YES NO

Appropriate declared pregnant woman procedures were implemented, and records of embryo/fetus dose were kept [1.7.8] YES NO NA

The registrant's radiation protection program incorporates considerations for keeping doses ALARA YES NO

Monitored individuals are properly presented with radiation exposure data at least annually [2.6] YES NO

Monitored individuals who have terminated employment were properly notified [2.6] YES NO NA

¹ 20=Dental Intraoral; 21= Dental Extraoral; 22= Cephalometric; 23=Panoramic; 24= Cone Beam CT (CBCT)

4) POSTING AND LABELING

- Agency Form RCA-1 ["Notice to Employees"] is properly posted [2.4(A)] YES NO
- RCA regulations and Certificate of Registration are posted, or a notice is posted stating where these documents are located [2.6] YES NO
- Operating procedures (including written radiation safety procedures) are available [2.6] YES NO
- Statement of deficiencies posted when applicable [2.6] YES NO NA
- Latex glove warning sign posted when applicable [2.5(B)] YES NO NA

5) REGISTRATION

- Equipment installation or servicing were obtained only from a registered Provider of X-Ray Services [3.3(A)] – PXS Registration #: YES NO
- Radiation physics services were obtained from only for a registered Radiation Physics Service provider [3.5.3] – RPS Registration #: YES NO NA
- Written notification was provided to the RCA before making any changes (e.g. installation of removal of x-ray systems) which would render the information contained in the Application for Registration inaccurate [3.8] YES NO NA
- For new registrations or facilities with modifications subsequent to previous inspection, a floor plan, shielding specification, and equipment arrangement have been submitted to the RCA prior to construction. [3.5.1(B)] YES NO NA
- For new registrations or facilities with modifications subsequent to previous inspection, the shielding plan was reviewed and evaluated by a registered Radiation Physics Service provider prior to X-Ray system use [3.5.1(C)] – RPS Registration #: YES NO NA

6) TECHNIQUE PROTOCOLS

- Type and size of the image receptor combination to be used, if any [4.3.4] YES NO
- Source to image receptor distance to be used, if any (typically present on the x-ray tube head) [4.3.4] YES NO
- Type and location of placement of patient shielding [4.3.6(A)(2)] YES NO

7) MAINTENANCE RECORDS & ASSOCIATED INFORMATION

Registrant maintains the following information for each X-Ray system (in a separate file or in chronological order):

[4.3.13(A)]

- | | | | | | | |
|--|--------------------------|------------|--------------------------|-----------|--------------------------|-----------|
| Maximum rating of technique factors (if applicable)
[4.3.13(A)(1)] | <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | <input type="checkbox"/> | NA |
| Model and serial numbers of all major components, and users' manuals for those components [4.3.13(A)(2)] | <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
| Aluminum equivalent filtration in the useful beam, including any routine variation (if applicable) [4.3.13(A)(3)] | <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
| Tube rating charts and cooling curves (if applicable)
[4.3.13(A)(4)] | <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | <input type="checkbox"/> | NA |
| Records of surveys, calibrations, maintenance & modifications performed on X-Ray systems and names of persons who performed such services [4.3.13(A)(5)] | <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
| Copy of correspondence with RCA regarding each X-ray system [4.3.13(A)(7)] | <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
-

8) X-RAY UTILIZATION LOG

The log includes the patient's name, the type of examinations, and the dates the examinations were performed [4.3.14(A)]

The log includes the name of the dentist who ordered the examination [4.3.14(A)(1)]

The log includes name(s) of individual(s) who performed the examination [4.3.14(A)(2)]

The log includes any deviation from the standard procedure as specified on the technique chart, including all repeat exposures [4.3.14(A)(3)]

When applicable, log includes X-Ray system used [4.3.14(A)(4)]

If human auxiliary support is required, log includes name of the human holder [4.3.14(A)(5)]

X-Ray utilization logs are maintained for a minimum of 5 years following the examination or treatment of adults and 5 years past the age of maturity for children [4.3.14(B)]

- | | | | | | |
|--------------------------|------------|--------------------------|-----------|--------------------------|-----------|
| <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
| <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
| <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
| <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | <input type="checkbox"/> | NA |
| <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | <input type="checkbox"/> | NA |
| <input type="checkbox"/> | YES | <input type="checkbox"/> | NO | | |
-

9) GENERAL REQUIREMENTS FOR DENTAL X-RAY

Label with the following is present on the control panel:
"WARNING: This X-Ray unit may be dangerous to patient and operator unless safe exposure factors and operating instructions and maintenance schedules are observed" [4.14.2(B)] **YES** **NO**

Where ≥ 2 radiographic tubes controlled by 1 exposure switch, tube(s) selected shall be clearly indicated prior to initiation of exposure and only selected tube can be energized [4.14.7] **YES** **NO** **NA**

Tube housing assembly supports adjusted, so assembly remains stable during an exposure (unless tube housing movement is designed function) [4.14.8] **YES** **NO**

All position locking, holding, and centering devices on X-Ray system components are functioning as intended [4.14.10] **YES** **NO**

For X-ray equipment capable of displaying technique factors, the technique factors used during exposure are indicated before exposure begins. If automatic exposure controls used, technique factors set prior to exposure are indicated [4.14.11(A)] **YES** **NO**

Means provided to terminate the exposure at a preset time interval, a preset product of current and time, a preset number of pulses, or a preset radiation exposure to the image receptor [4.14.13] **YES** **NO**

The half value layer meets the requirements set forth in Tables I and/or II [4.14.16] **YES** **NO** **NA**

10) SPECIAL REQUIREMENTS FOR DENTAL X-RAY

Facility utilizes intraoral dental radiographic unit(s) designed to be operated as a hand-held device [4.13(A)] **YES** **NO**

If **YES**, facility maintains documentation that each operator of hand-held X-ray system has completed training as specified by the manufacturer and approved by the Agency [4.13(A)] **YES** **NO** **NA**

If **YES**, unit is equipped with a backscatter shield of not less than 0.25 mm lead equivalent and 15.2 cm (6 inches) in diameter that is positioned as close as practicable to the distal end of the position indication device [4.13(A)] **YES** **NO** **NA**

If **YES**, facility follows protocols provided by the manufacturer, and approved by the Agency, regarding the safe operation of the unit [4.13(A)] **YES** **NO** **NA**

If **YES**, unit is secured from unauthorized removal or use [4.13(A)] **YES** **NO** **NA**

10) SPECIAL REQUIREMENTS FOR DENTAL X-RAY (cont.)Facility utilizes cone beam CT (CBCT) devices [4.7.8] YES NOIf YES, CBCT is only operated by an individual who has been specifically trained in its operation [4.7.8(D)] YES NO NA

11) QUALITY ASSURANCE PROGRAMQMP performed review of Quality Assurance Program at an interval not to exceed 12 months and provided a written report YES NO

Registrant has established and maintained a quality assurance program including:

Written standard operating procedures on radiation protection are reviewed and updated annually by management [4.10.1(A)(1)(a)] YES NOEmployee review and written acknowledgement of standard operating procedures and policies on radiation protection [4.10.1(A)(1)(b)] YES NODocumentation of minimum qualifications for dentists and X-Ray equipment operators [4.10.1(A)(1)(c)] YES NORecord retention in accordance with applicable Rhode Island statutes and regulations, but in no case less than 3 years [4.10.1(A)(1)(d)] YES NOCompliance with QA for image processing equipment [4.10.1(A)(2)(c)] YES NO

Radiographic Equipment:

System(s) evaluated by QMP prior to initial clinical use and after installation or relocation. Evaluation follows nationally recognized procedures or those recognized by the Agency [4.10.1(A)(3)(b)] YES NO

CBCT Equipment:

Evaluation of CBCT performed by QMP within 30 days of initial installation, at intervals not to exceed 12 months, and within 30 days after any change or replacement of components which, in the opinion of the QMP, could cause a change in the radiation output or image quality [4.10.1(A)(5)(e)] YES NO NA

Inspection Report #:

Registration #: DEF-xxxx

12) GENERIC COMMUNICATIONS

RCA Information Notices, Newsletters and other generic communications are being received

YES NO NA

When required, appropriate training and action is being taken in response to these generic communications

YES NO NA

13) EXIT MEETING

SAMPLE

APPENDIX A - DOCUMENTATION OF DEFICIENCIES

Reference		Basis for Deficiency
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		
Report Item:		
Regulation:		
Type N/C:		

APPENDIX B - ACTION ON PREVIOUS INSPECTION FINDINGS

Identification and Summary of Action Taken

Status

Report No.:

Type N/C:

OPEN

CLOSED

Description of Deficiency:	
Action Taken:	

SAMPLE