COMPANY NAME_____ **ADDRESS**_____

THERMOMETER ACCURACY LOG

Procedure: The stem thermometer will be placed into an ice slurry for approximately 1 minute. The thermometer will be adjusted to within 2° of 32°, if necessary. The cooler thermometer(s) will be calibrated by placing the calibrated stem in the cooler near the probe. The temperature readings will be written below.

DATE	STEM	COOLER	INITIAL	VERIFIED BY	DATE